








View A







T= Load-bearing seam (Tragende Naht)
N= Non-load-bearing seam (Nichttragende Naht)
D= Seal seam (Dichtnaht)

ZfP-Prüfverfahren / test procedure:

VT = Visual Testing (Visuelle Prüfung)
RT = Radiographfic Testing according to AD2000 HP0/HP100R
(Durchstrahlungsprüfung)
LT = Leakage Testing (Leckageprüfung)

Helium tight welding
Helium tight at 20 bar (Design pressure)
Operating temperature 4K (-269°C)
Design and manufacturing of pressure equipment according to
Pressure Vessel Regulations AD 2000 , specification sheet HP 0

Position 1 stretched length: ~77
Position 2 stretched length: ~202
Alternative material (Pos.1&2):
1.4306, 1.4541, 1.4311, 1.4335, 1.4429,
1.4435, 1.4436, 1.4438, 1,4466, 1.4563

(1) According to DIN EN 10216-5 / EN ISO 1127 D4/T3
Acceptance test certificate according to DIN EN 10204/3.1
Acceptance according to AD 2000-W2/W10

Welding classification DIN EN ISO 5817	Quality level
Helium weldings	-B-

Edges ISO 13715 +0.3 -0.3 -0.1 Rz 25

Copying of this document and giving it to others and the use of communication of the contents thereof are forbidden without express authority. Offenders are liable to the payment of damages. All rights are reserved in the event of the grant of a patent or the registration of a utility model or design.

Projection

Change Number

Surfaces
DIN EN ISO 1302:2002

----Approved----

GSI
Helmholtzzentrum für Schwerionenforschung GmbH
Planckstrasse 1, 64291 Darmstadt

Dimensions without tolerance indication
DIN ISO 2768-1-m
DIN ISO 2768-2-K:1991
DIN EN ISO 13920-BF

Weight 0 kg

Material

Scale 1:1

	Date	Name
Drawn	29.09.2025	PDOLL
Checked	06.10.2025	VETTER
Approved	06.10.2025	ZURKAN

Description: Helium pipe out - VC

Size A2

Drawing No. MG-1552864-A-000

Rev. -

Sheet 1 of 1

FR