



NOTES:

1. LEAK RATE $< 1 \times 10^{-10}$ mbar l/s
2. MATERIAL CERTIFICATES ARE REQUIRED
3. MAGNETIC PERMEABILITY $\mu_r \leq 1.05$
4. OUTGASSING RATE $< 5 \times 10^{-10}$ mbar l/s cm²

WLB-23130
FLT-23130-1
VLC-3
SLT-3
DCC-2
CLN-2
VLC-2
PKG-23130

POS.	DESCRIZIONE	Q.TA'	MATERIALE	DIM. GREZZE/COD.
3	SPECIAL WELD FLANGE DN160CF ROTATABLE	1	AISI 316L	23130-1-14
2	SPECIAL WELD FLANGE DN160CF FIXED	1	AISI 316L	23130-3-12A
1	SPECIAL BELLOWS WITH ENDS	1	AISI 316L	23130-3-10

TOLLERANZE GENERALI	< 6	6-30	30-120	120-400	400-1000	>1000	< 10	10-50	50-120	120-400	> 400	
General tolerances	UNI EN ISO 22768 - fH <input type="checkbox"/>	±0.05	±0.1	±0.15	±0.2	±0.3	±0.5	±1°	±0°30'	±0°20'	±0°10'	±0°5'
	UNI EN ISO 22768 - mK <input checked="" type="checkbox"/>	±0.1	±0.2	±0.3	±0.5	±0.8	±1.2					

A TERMINI DI LEGGE CI RISERVIAMO LA PROPRIETA' DI QUESTO DISEGNO CON DIVIETO DI RIPRODURLO O DI RENDERLO COMUNE NOTO A TERZI SENZA LA NOSTRA AUTORIZZAZIONE

AS BY LAW ENACTED WE RESERVE THE RIGHT OF OWNERSHIP OF DRAWING. THE DOCUMENT AND INFORMATION CONTAINED HEREIN MAY NOT BE COPIED USED OR DISCLOSED WITHOUT OUR PERMISSION

SPIGOLI NON QUOTATI ISO 13715
Edges of undefined shape ISO 13715

SCALA Scale **A3**

FORMATO Size **A3**

DESIGNATO Draw	DANIELE EBRI	08/11/2023
VERIFICATO Controlled	MARCO CANETTI	08/11/2023
APPROVATO Released	ANTONIO CORSI	08/11/2023

DESCRIZIONE Description	SPECIAL BELLOWS ASSEMBLIES	23130-3-00	REVISIONE Revision	A
----------------------------	-----------------------------------	-------------------	-----------------------	----------