



① fitted to outside diameter item 2 at the choice of manufacturer

Integral leak rate $< 1 \times 10^{-10}$ mbar l/s

Outgassing rate $< 5 \times 10^{-10}$ mbar l/s cm²

pay attention to F-TG-QUA-de-Schweisstechnik-V007

UHV-suitable cleaning, an example can be found in the guideline F-TG-V6.1e and F-TG-V-6.3e

Details, also to the alternative materials can be found in F-DS-HEB-en-VC_0096_Detailed_Specification_Straight_Tubes

Welding classification	Quality level
DIN EN ISO 5817	
Vacuum weldings	-B-

Edges ISO 13715 $\begin{matrix} +0.3 \\ -0.1 \end{matrix}$ $\sqrt{Rz25}$

<small>Copying of this document and giving it to others and the use of communication of the contents thereof are forbidden without express authority. Offenders are liable to the payment of damages. All rights are reserved in the event of the grant of a patent or the registration of a utility model or design.</small>		Projection 	Change Number -----Approved-----
GSF <small>Heilmittelzentrum für Schwerionenforschung GmbH Planckstrasse 1, 64291 Darmstadt</small>		Dimensions without tolerance indication <small>DIN ISO 2768-1-1 DIN ISO 2768-2-1:1991 DIN EN ISO 1302:2002</small>	Weight 30,1 kg Scale 1:1
Drawn 03.07.2025 EHEGER Checked 07.08.2025 ZURKAN Approved 22.09.2025 TEBERL	Date Name Description Beamtube	Drawing No. VC-1450932-A-V02 Rev. -	Size A1 Sheet 1 of 1